



Physikalisch-Technische Bundesanstalt
Braunschweig und Berlin

OIML Member State

Germany

OIML Certificate No.

R60/2000-A-DE1-2022.01

OIML CERTIFICATE ISSUED UNDER SCHEME A

OIML Issuing Authority

Name: Physikalisch-Technische Bundesanstalt,
Conformity Assessment Body
Address: Bundesallee 100, 38116 Braunschweig, GERMANY
Person responsible: Dr.-Ing Prof. h.c. Frank Härtig

Applicant

Name: BROSA GmbH
Address: Dr.-Klein-Str. 1, 88069 Tettngang, Germany

Manufacturer

Name: BROSA GmbH
Address: Dr.-Klein-Str. 1, 88069 Tettngang, Germany

Identification of the certified type (*the detailed characteristics will be defined in the additional pages*)

Load cell
Type: 0201

Designation of the module (*if applicable*)

Load cell equipped with electronics

This OIML Certificate attests the conformity of the above identified type (represented by the sample(s) identified in the OIML type evaluation report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 60

Edition (year): 2000

For accuracy class (if applicable): D0.1

**OIML Certificate No.
R60/2000-A-DE1-2022.01**

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report:

No. 1.12-4095016 dated 31.03.2022 that includes 6 pages

The technical documentation relating to the identified type is contained in documentation file:

No. ZDS-R60/2000-A-DE1-2022.01 dated 28.04.2022 that includes 2 pages

OIML Certificate History

Revision No.	Date	Description of the modification
-	28.04.2022	Initial issuance

Identification, signature and stamp

The Issuing Authority


Dr. Oliver Mack

Member of Conformity Assessment Body

Date: 28.04.2022



Table 1: Essential data

Accuracy class		D0.1
Max. number of load cell intervals	n_{LC}	100
Maximum capacity	E_{max}	t 16
Minimum load cell verification interval	$V_{min} = (E_{max} / Y)$	$E_{max} / 500$

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

