

**OIML Certificate No.
R76/2006-A-DE1-2022.01**

This OIML Certificate relates only to metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML Recommendation identified above.

This OIML Certificate does not bestow any form of legal international approval.

The conformity was established by the results of tests and examinations provided in the associated OIML type evaluation report:

No. 1.12-4095017 dated 27.01.2022 that includes 7 pages

The technical documentation relating to the identified type is contained in documentation file:

No. ZDS-R76/2006-A-DE1-2022.01 dated 18.02.2022 that includes 2 pages

OIML Certificate History

Revision No.	Date	Description of the modification
First issuance	18.02.2022	---

Identification, signature and stamp

The Issuing Authority


Dr. Oliver Mack

Member of Conformity Assessment Body

Date: 18.02.2022



Table 1: Essential Data

Accuracy class	III, IIII
Weighing range	Single interval
Power supply	10 - 30 V DC
Maximum number of scale intervals	10000
Fraction of the maximum permissible error	0

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate is issued, partial quotation of the Certificate and of the associated OIML type evaluation report(s) is not permitted, although either may be reproduced in full.

